

(Time: 3 hours)

Total Marks: 100

- N.B.: (1) All questions are compulsory.
 (2) Figures to the right indicate full marks.
 (3) Use of log table/ non-programmable calculator is allowed.

Q.1 Attempt any four of the following. 20

- A) Discuss various grades of laboratory reagents.
- B) Calculate mass percent composition of each element in $\text{CH}_3\text{CH}_2\text{OH}$, ethanol molecule.
 (Given: atomic mass of C = 12, H = 1, O = 16)
- C) Calculate the number of grams of pure Na_2CO_3 required to prepare 250 cm^3 of 0.1 N solution. This Na_2CO_3 is to be titrated with HCl according to equation $\text{CO}_3^{2-} + 2 \text{H}^+ \rightarrow \text{H}_2\text{CO}_3$
 (Given: atomic mass of Na = 23, C = 12, O = 16)
- D) Describe the sampling of homogeneous and heterogeneous liquid.
- E) Name different methods of reduction of sample size in sampling of solid. Explain any two of them.
- F) Explain the terms quality and quality control.

Q.2 Attempt any four of the following. 20

- A) What are the advantages and limitations of EDTA as titrant?
- B) What are redox indicators? Explain the use of iron(II) orthophenanthroline indicator.
- C) Calculate the potential of the system in the titration of 10 cm^3 of 0.2M Fe(II) against 0.2M Ce(IV) solution when:
 i) 12 cm^3 Ce(IV) solution added.
 ii) 15 cm^3 Ce(IV) solution added.
 ($E^\circ_{\text{Pt}|\text{Fe}^{2+},\text{Fe}^{3+}} = +0.771 \text{ V}$, $E^\circ_{\text{Pt}|\text{Ce}^{3+},\text{Ce}^{4+}} = +1.440 \text{ V}$)
- D) How is the selectivity is enhanced in the complexometric titrations by the following:
 i) kinetic masking ii) use of masking and demasking agents.
- E) What are metal-ion indicators? What are their requirements?
- F) Derive an equation for potential of the system at equivalence point in the titration of Fe(II) against MnO_4^- .

Q.3. Attempt any four of the following. 20

- A) Discuss the principle of Atomic Absorption Spectroscopy. Explain the role of rotating chopper in AAS.
- B) Discuss the basic principles of FES.
- C) Explain the applications of fluorescence spectroscopy.

- D) Explain the principles underlying the emission of fluorescent light. How does fluorescence differ from phosphorescence?
- E) Draw a schematic diagram of turbidimeter and explain turbidimetric titrations using turbidimetric titration curve.
- F) What are the important factors affecting the scattering of radiation? Explain any two.

Q.4 Attempt any four of the following. 20

- A) Explain any two factors affecting solvent extraction.
- B) Discuss the different steps involved in process of solid phase extraction.
- C) Explain the terms used in HPLC i) isocratic elution ii) gradient elution. Give any three applications of HPLC.
- D) With a neat and labelled diagram explain the working of HPLC.
- E) Give applications and limitations of HPTLC.
- F) Explain double beam densitometer used in HPTLC. Give any two advantages of HPTLC.

Q.5 A) Select the correct option and complete the following statements: (any five) 05

- a) _____ material can be used for verification of method validation parameters.
i) Reference material ii) Certified reference material iii) LR grade
- b) The next step after quality control is _____.
i) quality management ii) quality assurance iii) quality development
- c) The sum of mole fraction of solute and solvent is _____.
i) 1 ii) 1.5 iii) 2.5
- d) _____ is a method of expressing concentration in a solution on weight basis.
i) Normality ii) Molality iii) Molarity
- e) Flushing method is used for sampling of _____.
i) Solid ii) liquid iii) gases
- f) _____ is used for sampling of compact solid.
i) Auger sampler ii) Multiple tube sampler iii) Split tube thief
- g) The ratio of weight of sample to total weight of bulk is _____.
i) Bulk size ii) Size:weight ratio iii) Bulk ratio
- h) _____ is used for sampling of flowing liquid.
i) Concentric tube thief ii) Geo-sampler iii) multiple tube sampler

Q.5 B) State whether true or false: (any five) 05

- a) In the titration of iron(II) with cerium(IV); iron(II) undergoes reduction.
- b) Potassium permanganate acts as self indicator in redox titrations.
- c) Nernst's distribution law is applicable in redox titrations.
- d) pCa is defined as $\log_{10}[\text{Ca}^{+2}]$
- e) Diphenyl amine is the first indicator to be used in redox titrations.

- f) Dissociation of EDTA is not affected by pH.
- g) Murexide is obtained from purpuric acid.

Q.5 C) Fill in the blanks with correct alternatives given in the bracket: 05
(any five)

(acetylene, phosphorescence, turbidity coefficient, turbidance, toxic metals, mist, low, right)

- a) AAS is used to detect _____ like Cu, Ni, Zn and Hg in food products.
- b) The fuel used in a premix burner in FES is _____.
- c) The nebulizer converts the sample solution into a _____.
- d) Delayed re-emission of absorbed radiation is called _____.
- e) Phosphorimetric experiments are normally carried out at _____ temperature.
- f) In the expression w.r.t. turbidimetry, $S = Ktc$, 'K' stands for _____.
- g) In nephelometry, the detector is usually, but not necessarily, placed at _____ angle to the incident radiation.
- h) A turbidimeter measures _____ as a function of concentration of suspensions.

Q.5 D) Match the columns: (any five) 05

- | Column A | Column B |
|---|---|
| a) TBP solvent | (i) Ion-pair formation |
| b) Multistage separation | (ii) Easily extracted in organic solvent |
| c) Hydrophobic functional groups chemically bonded silica | (iii) Analytical column |
| d) Neutral chelate | (iv) HPTLC |
| e) Complexes form clusters | (v) Solid phase extraction |
| f) HPLC | (vi) $pH_{1/2}$ |
| g) Densitometer | (vii) Easily extracted in aqueous solvent |
| | (viii) Countercurrent extraction |
| | (ix) Extraction of uranyl nitrate |
